

Questions from tenderers and corresponding answers

N.B.: The questions and answers will form an integral part of the tender documents.

Date: 08 Jan, 2020

Requested by: ICEM Asia For the project “Mekong River Commission – Environmental Monitoring Pilots in Lao PDR”

Funded by GIZ - transaction number 81245468

Ref.: AAS1905-RFQ-001

Supply and Delivery of: Monitoring equipment (including installation and training, where applicable)

No.	Question	Answer
1	Do you require a trailer for the boat?	Yes
2	Is the boat only required to be delivered to Vientiane?	Yes, the boat should be delivered to the Vientiane address included in the RFQ
3	Does the boat require a full roof with enclosure, or a part roof and open sides?	The boat should have open sides
4	Should the boats steering be located at the rear of the boat or in the mid -section?	Mid-section steering is preferable
5	Should the boat engine be 2-stroke or 4-stroke?	The boat engine should be 2-stroke
6	For the logging WQ meter – should the cable lengths as specified be between the probe and the logging unit, or so that the logging unit can be lowered into the water?	The specified cable lengths are to be between the probe and logging unit, so that the probe can be lowered to the appropriate depth in the water.
7	How would the system be deployed? Will the logger and the sensors be deployed left underwater while it is collecting data? Or would it be deployed from a boat to collect data and then retrieved back?	The system is for manual logging, and must be able to be deployed from the river bank or from a boat.
8	For the portable field radiometer - should the cable lengths as specified be between the probe and the logging unit, or between the logging unit and a computer?	The specified cable lengths are to be between the probe and logging unit, so that the probe can be lowered to the appropriate depth in the water.
9	For billing for equipment procurement – Who should the bills be invoiced to?	Invoice(s) will be billed to the below company name and address: ICEM Asia Consulting Pte. Ltd – GIZ project Address: 60 Paya Lebar Road #11-01 Paya Lebar square, 409051 SINGAPORE